

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062223\
 Data File : VN078348.D
 Acq On : 22 Jun 2023 14:28
 Operator : JC\MD
 Sample : 03289-13
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-05-12.0-14.0

Quant Time: Jun 22 17:02:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	439252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	781752	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	719203	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	257647	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	293841	50.173	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.340%
35) Dibromofluoromethane	8.171	113	271838	54.995	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.000%
50) Toluene-d8	10.571	98	908228	47.662	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.320%
62) 4-Bromofluorobenzene	12.853	95	311701	52.618	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062223\
Data File : VN078348.D
Acq On : 22 Jun 2023 14:28
Operator : JC\MD
Sample : 03289-13
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-05-12.0-14.0

Quant Time: Jun 22 17:02:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

