

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062223\
 Data File : VN078350.D
 Acq On : 22 Jun 2023 15:16
 Operator : JC\MD
 Sample : 03289-23
 Misc : 5.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-08-4.0-6.0

Quant Time: Jun 22 17:02:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	490554	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	9.106	114	878968	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	809660	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	299767	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	326707	49.951	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.900%
35) Dibromofluoromethane	8.171	113	304075	54.713	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.420%
50) Toluene-d8	10.571	98	1010918	47.184	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.360%
62) 4-Bromofluorobenzene	12.853	95	356619	53.542	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062223\
Data File : VN078350.D
Acq On : 22 Jun 2023 15:16
Operator : JC\MD
Sample : 03289-23
Misc : 5.10g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-08-4.0-6.0

Quant Time: Jun 22 17:02:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

