

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062223\
 Data File : VN078352.D
 Acq On : 22 Jun 2023 16:04
 Operator : JC\MD
 Sample : 03378-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB-01-062123

Quant Time: Jun 23 08:02:50 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.236	168	476670	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	864484	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	754997	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	231909	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	320095	50.365	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.740%	
35) Dibromofluoromethane	8.177	113	303598	55.543	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 111.080%	
50) Toluene-d8	10.571	98	979196	46.469	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 92.940%	
62) 4-Bromofluorobenzene	12.853	95	311417	47.539	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.080%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	5.289	84	10562	1.747	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062223\
Data File : VN078352.D
Acq On : 22 Jun 2023 16:04
Operator : JC\MD
Sample : 03378-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB-01-062123

Quant Time: Jun 23 08:02:50 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

