

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062223\
 Data File : VN078354.D
 Acq On : 22 Jun 2023 16:52
 Operator : JC\MD
 Sample : 03378-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GWOW-BR-05-345-365-062123

Quant Time: Jun 23 08:03:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	506344	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	933534	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	823462	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	257958	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	344632	51.048	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.100%	
35) Dibromofluoromethane	8.177	113	323626	54.828	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.660%	
50) Toluene-d8	10.571	98	1084495	47.659	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.320%	
62) 4-Bromofluorobenzene	12.853	95	341301	48.247	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.500%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	2.524	62	5394	0.720	ug/l	90
12) 1,1-Dichloroethene	4.353	96	815514	147.490	ug/l	92
21) trans-1,2-Dichloroethene	5.794	96	60304	10.112	ug/l	91
24) 1,1-Dichloroethane	6.583	63	74233	7.088	ug/l	99
27) cis-1,2-Dichloroethene	7.500	96	384312	54.025	ug/l	95
40) Benzene	8.624	78	7843	0.310	ug/l	94
42) 1,2-Dichloroethane	8.677	62	43183	4.983	ug/l	99
44) Trichloroethene	9.359	130	397310	59.520	ug/l	94
49) 1,4-Dioxane	9.700	88	10234	91.728	ug/l #	93
52) Toluene	10.635	92	56299	3.528	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062223\
Data File : VN078354.D
Acq On : 22 Jun 2023 16:52
Operator : JC\MD
Sample : 03378-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GWOW-BR-05-345-365-062123

Quant Time: Jun 23 08:03:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

