

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062223\
 Data File : VN078356.D
 Acq On : 22 Jun 2023 17:40
 Operator : JC\MD
 Sample : 03360-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CE-91-062023

Quant Time: Jun 23 08:03:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	461907	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	872056	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	809707	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	248593	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	322767	52.409	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.820%	
35) Dibromofluoromethane	8.177	113	295590	53.608	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.220%	
50) Toluene-d8	10.571	98	1044530	49.139	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.280%	
62) 4-Bromofluorobenzene	12.853	95	328360	49.690	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 99.380%	
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	79853	45.046	ug/l	100
20) Methylene Chloride	5.283	84	25228	4.307	ug/l	92
25) 2-Butanone	7.494	43	20151	7.392	ug/l	99
37) Ethyl Acetate	7.571	43	54344	8.796	ug/l	97
43) Isopropyl Acetate	8.694	43	276781	23.424	ug/l #	83
51) 4-Methyl-2-Pentanone	10.447	43	147526	23.709	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062223\
Data File : VN078356.D
Acq On : 22 Jun 2023 17:40
Operator : JC\MD
Sample : 03360-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CE-91-062023

Quant Time: Jun 23 08:03:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

