

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062320\
 Data File : VN062043.D
 Acq On : 23 Jun 2020 15:33
 Operator : JC/MD
 Sample : VN0623WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampled :
 VN0623WBL01

Quant Time: Jun 24 01:47:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	155242	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	288048	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	276507	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	103059	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.98	65	108359	52.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.10%	
35) Dibromofluoromethane	7.54	113	90288	49.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.00%	
50) Toluene-d8	10.06	98	361045	49.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.68%	
62) 4-Bromofluorobenzene	12.38	95	116927	47.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.10%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062320\
Data File : VN062043.D
Acq On : 23 Jun 2020 15:33
Operator : JC/MD
Sample : VN0623WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0623WBL01

Quant Time: Jun 24 01:47:48 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

