

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062320\
 Data File : VN062046.D
 Acq On : 23 Jun 2020 17:00
 Operator : JC/MD
 Sample : L3043-08DL 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE122D1-20200615DL

Quant Time: Jun 24 01:55:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	140124	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	254973	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	248067	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	103596	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	96127	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
35) Dibromofluoromethane	7.54	113	80712	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
50) Toluene-d8	10.06	98	323955	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
62) 4-Bromofluorobenzene	12.38	95	111103	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
Target Compounds						
44) Trichloroethene	8.80	130	141394	76.263	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062320\
Data File : VN062046.D
Acq On : 23 Jun 2020 17:00
Operator : JC/MD
Sample : L3043-08DL 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE122D1-20200615DL

Quant Time: Jun 24 01:55:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

