

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062320\
 Data File : VN062047.D
 Acq On : 23 Jun 2020 17:25
 Operator : JC/MD
 Sample : L3043-10DL 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE139D2-20200616DL

Quant Time: Jun 24 01:57:34 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	140447	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	260533	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	249871	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	97135	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	98637	52.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.74%	
35) Dibromofluoromethane	7.55	113	81911	49.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.28%	
50) Toluene-d8	10.06	98	330339	50.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.82%	
62) 4-Bromofluorobenzene	12.38	95	111000	49.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.76%	
Target Compounds						
44) Trichloroethene	8.80	130	59088	31.190	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062320\
Data File : VN062047.D
Acq On : 23 Jun 2020 17:25
Operator : JC/MD
Sample : L3043-10DL 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE139D2-20200616DL

Quant Time: Jun 24 01:57:34 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

