

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062320\
 Data File : VN062048.D
 Acq On : 23 Jun 2020 17:49
 Operator : JC/MD
 Sample : L3043-04DL 10X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE103D1-20200616DL

Quant Time: Jun 24 02:00:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	144956	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	270676	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	267216	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	104784	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	101833	52.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.76%	
35) Dibromofluoromethane	7.55	113	85225	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
50) Toluene-d8	10.06	98	344156	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
62) 4-Bromofluorobenzene	12.38	95	119087	51.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.00%	
Target Compounds						
44) Trichloroethene	8.80	130	117175	59.534	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062320\
Data File : VN062048.D
Acq On : 23 Jun 2020 17:49
Operator : JC/MD
Sample : L3043-04DL 10X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE103D1-20200616DL

Quant Time: Jun 24 02:00:15 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

