

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062320\
 Data File : VN062049.D
 Acq On : 23 Jun 2020 18:13
 Operator : JC/MD
 Sample : L3043-07DL 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT101D2-20200615DL

Quant Time: Jun 24 02:02:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 17 01:19:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	149422	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	276719	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	266220	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	97249	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	105102	52.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.90%	
35) Dibromofluoromethane	7.55	113	88043	49.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.46%	
50) Toluene-d8	10.06	98	349282	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
62) 4-Bromofluorobenzene	12.38	95	111407	46.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.34%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	3.71	101	2056	1.357	ug/l	94
44) Trichloroethene	8.80	130	142013	70.577	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062320\
Data File : VN062049.D
Acq On : 23 Jun 2020 18:13
Operator : JC/MD
Sample : L3043-07DL 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT101D2-20200615DL

Quant Time: Jun 24 02:02:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061620W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 17 01:19:44 2020
Response via : Initial Calibration

