

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062321\
 Data File : VN067451.D
 Acq On : 23 Jun 2021 11:13
 Operator : JC/MD
 Sample : VN0623MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0623MBL01

Quant Time: Jun 24 05:02:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	267490	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	556936	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	481439	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	151269	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	252486	47.85	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.70%
35) Dibromofluoromethane	8.024	113	170600	46.06	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.12%
50) Toluene-d8	10.443	98	735445	51.80	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.60%
62) 4-Bromofluorobenzene	12.733	95	240396	41.03	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	82.06%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062321\
Data File : VN067451.D
Acq On : 23 Jun 2021 11:13
Operator : JC/MD
Sample : VN0623MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0623MBL01

Quant Time: Jun 24 05:02:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

