

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062321\
 Data File : VN067462.D
 Acq On : 23 Jun 2021 17:16
 Operator : JC/MD
 Sample : M2804-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW181S-20210621

Manual Integrations
 APPROVED

MMDadoda
 6/24/2021 4:59:04 PM

Quant Time: Jun 24 05:04:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	273336	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	574334	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	515523	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	172849	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	258361	47.91	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	95.82%		
35) Dibromofluoromethane	8.024	113	172615	45.19	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	90.38%		
50) Toluene-d8	10.443	98	772302	52.75	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	105.50%		
62) 4-Bromofluorobenzene	12.733	95	254959	42.20	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	84.40%		
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	7.318	96	6001m	1.40	ug/l	
30) Chloroform	7.825	83	5595m	0.71	ug/l	
32) 1,1,1-Trichloroethane	8.027	97	3006	0.41	ug/l #	52
44) Trichloroethene	9.212	130	7537m	1.79	ug/l	
64) Tetrachloroethene	10.979	164	4316	1.18	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062321\
Data File : VN067462.D
Acq On : 23 Jun 2021 17:16
Operator : JC/MD
Sample : M2804-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW181S-20210621

Manual Integrations
APPROVED

MMDadoda
6/24/2021 4:59:04 PM

Quant Time: Jun 24 05:04:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

