

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN062321\  
 Data File : VN067465.D  
 Acq On : 23 Jun 2021 18:29  
 Operator : JC/MD  
 Sample : M2804-05 10X  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 DUP08-20210621

Quant Time: Jun 24 05:04:40 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N061421W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jun 15 05:19:42 2021  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 8.088  | 168   | 266388   | 50.00    | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 8.971  | 114   | 560166   | 50.00    | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 11.747 | 117   | 493245   | 50.00    | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.678 | 152   | 155496   | 50.00    | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 8.440  | 65    | 255570   | 48.63    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =     | 97.26%   |
| 35) Dibromofluoromethane    | 8.024  | 113   | 170105   | 45.66    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =     | 91.32%   |
| 50) Toluene-d8              | 10.446 | 98    | 741924   | 51.96    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =     | 103.92%  |
| 62) 4-Bromofluorobenzene    | 12.734 | 95    | 247588   | 42.01    | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =     | 84.02%   |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN062321\  
Data File : VN067465.D  
Acq On : 23 Jun 2021 18:29  
Operator : JC/MD  
Sample : M2804-05 10X  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
DUP08-20210621

Quant Time: Jun 24 05:04:40 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N061421W.M  
Quant Title : SW846 8260  
QLast Update : Tue Jun 15 05:19:42 2021  
Response via : Initial Calibration

