

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062321\
 Data File : VN067466.D
 Acq On : 23 Jun 2021 18:53
 Operator : JC/MD
 Sample : M2804-06 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP07-20210621

Quant Time: Jun 24 05:04:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 15 05:19:42 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	259606	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	554830	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	491091	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	167104	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	250004	48.82	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.64%
35) Dibromofluoromethane	8.024	113	166606	45.15	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.30%
50) Toluene-d8	10.443	98	741561	52.43	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.86%
62) 4-Bromofluorobenzene	12.734	95	249912	42.82	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	85.64%
Target Compounds						
44) Trichloroethene	9.223	130	9358	2.31	ug/l	Qvalue 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062321\
Data File : VN067466.D
Acq On : 23 Jun 2021 18:53
Operator : JC/MD
Sample : M2804-06 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP07-20210621

Quant Time: Jun 24 05:04:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 15 05:19:42 2021
Response via : Initial Calibration

