


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062321\
 Data File : VN067449.D
 Acq On : 23 Jun 2021 9:24
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061421W.M
 Title : SW846 8260
 Last Update : Tue Jun 15 05:19:42 2021



AutoFind: Scans 4005, 4006, 4007; Background Corrected with Scan 3988

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.9	18464	PASS
75	95	30	60	59.5	47893	PASS
95	95	100	100	100.0	80544	PASS
96	95	5	9	6.6	5356	PASS
173	174	0.00	2	1.5	693	PASS
174	95	50	100	57.4	46197	PASS
175	174	5	9	8.4	3870	PASS
176	174	95	101	95.1	43955	PASS
177	176	5	9	6.8	3006	PASS