

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062322\
 Data File : VN073038.D
 Acq On : 23 Jun 2022 10:41
 Operator : JC\MD
 Sample : VN0623WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0623WBL01

Quant Time: Jun 24 01:16:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.010	168	298406	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	481523	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	420512	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	177366	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	182042	45.123	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.240%
35) Dibromofluoromethane	7.951	113	154674	51.969	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.940%
50) Toluene-d8	10.375	98	495815	44.770	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.540%
62) 4-Bromofluorobenzene	12.669	95	195404	49.068	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062322\
Data File : VN073038.D
Acq On : 23 Jun 2022 10:41
Operator : JC\MD
Sample : VN0623WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0623WBL01

Quant Time: Jun 24 01:16:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

