

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062322\
 Data File : VN073057.D
 Acq On : 23 Jun 2022 18:42
 Operator : JC\MD
 Sample : N3431-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-14B-2-4

Quant Time: Jun 24 01:20:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	313127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	518607	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.681	117	468821	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	234590	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	194535	45.952	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.900%		
35) Dibromofluoromethane	7.951	113	163509	51.009	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	102.020%		
50) Toluene-d8	10.375	98	535031	44.857	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	89.720%		
62) 4-Bromofluorobenzene	12.669	95	235062	54.806	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	109.620%		
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	78722	33.829	ug/l	91
18) Methyl Acetate	4.793	43	7859	1.268	ug/l #	73
20) Methylene Chloride	5.016	84	28269	7.627	ug/l	90
25) 2-Butanone	7.269	43	22116	5.989	ug/l	89
31) Cyclohexane	8.016	56	14293	2.496	ug/l #	60
37) Ethyl Acetate	7.340	43	47333	6.823	ug/l	98
39) Methylcyclohexane	9.392	83	35128	6.719	ug/l	96
43) Isopropyl Acetate	8.493	43	209687	21.144	ug/l #	86
59) 2-Hexanone	11.028	43	245937	49.169	ug/l #	46
73) Isopropylbenzene	12.516	105	243633	13.424	ug/l	100
78) n-propylbenzene	12.857	91	185767	9.306	ug/l	97
83) tert-Butylbenzene	13.257	119	22637	1.658	ug/l	90
84) 1,2,4-Trimethylbenzene	13.304	105	17032	1.147	ug/l	85
85) sec-Butylbenzene	13.439	105	90679	4.979	ug/l	81
89) n-Butylbenzene	13.875	91	14212	1.180	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062322\
Data File : VN073057.D
Acq On : 23 Jun 2022 18:42
Operator : JC\MD
Sample : N3431-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-14B-2-4

Quant Time: Jun 24 01:20:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

