

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062322\
 Data File : VN073037.D
 Acq On : 23 Jun 2022 10:00
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 24 01:16:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	79	0.00
2 T	Dichlorodifluoromethane	0.476	0.521	-9.5	83	0.00
3 P	Chloromethane	0.535	0.509	4.9	81	0.00
4 C	Vinyl Chloride	0.427	0.470	-10.1#	85	0.00
5 T	Bromomethane	0.168	0.186	-10.7	80	0.00
6 T	Chloroethane	0.267	0.307	-15.0	91	0.00
7 T	Trichlorofluoromethane	0.695	0.880	-26.6#	108	0.01
8 T	Diethyl Ether	0.359	0.391	-8.9	85	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.481	0.506	-5.2	84	0.01
10 T	Methyl Iodide	0.551	0.357	35.2#	51	0.00
11 T	Tert butyl alcohol	0.175	0.174	0.6	77	-0.02
12 CM	1,1-Dichloroethene	0.479	0.496	-3.5#	81	0.00
13 T	Acrolein	0.069	0.061	11.6	70	0.00
14 T	Allyl chloride	1.023	1.068	-4.4	83	0.00
15 T	Acrylonitrile	0.403	0.439	-8.9	85	0.00
16 T	Acetone	0.372	0.391	-5.1	86	-0.01
17 T	Carbon Disulfide	1.227	1.104	10.0	72	0.00
18 T	Methyl Acetate	0.990	1.145	-15.7	96	0.00
19 T	Methyl tert-butyl Ether	1.798	2.089	-16.2	89	0.00
20 T	Methylene Chloride	0.628	0.632	-0.6	86	0.00
21 T	trans-1,2-Dichloroethene	0.521	0.541	-3.8	83	0.00
22 T	Diisopropyl ether	1.907	2.278	-19.5	92	0.00
23 T	Vinyl Acetate	1.760	2.010	-14.2	87	0.00
24 P	1,1-Dichloroethane	1.000	1.152	-15.2	90	-0.01
25 T	2-Butanone	0.590	0.619	-4.9	82	0.00
26 T	2,2-Dichloropropane	0.898	0.643	28.4#	59	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.699	-12.4	90	0.00
28 T	Bromochloromethane	0.413	0.467	-13.1	86	0.00
29 T	Tetrahydrofuran	0.389	0.416	-6.9	84	0.00
30 C	Chloroform	1.012	1.173	-15.9#	90	0.00
31 T	Cyclohexane	0.914	0.902	1.3	83	0.00
32 T	1,1,1-Trichloroethane	0.888	0.980	-10.4	85	0.00
33 S	1,2-Dichloroethane-d4	0.676	0.753	-11.4	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
35 S	Dibromofluoromethane	0.309	0.331	-7.1	86	0.00
36 T	1,1-Dichloropropene	0.418	0.450	-7.7	87	0.00
37 T	Ethyl Acetate	0.669	0.726	-8.5	89	-0.01
38 T	Carbon Tetrachloride	0.440	0.454	-3.2	83	0.00
39 T	Methylcyclohexane	0.504	0.495	1.8	78	0.00
40 TM	Benzene	1.314	1.425	-8.4	87	0.00
41 T	Methacrylonitrile	0.317	0.293	7.6	76	0.00
42 TM	1,2-Dichloroethane	0.467	0.531	-13.7	90	0.00
43 T	Isopropyl Acetate	0.956	1.059	-10.8	87	0.00
44 TM	Trichloroethene	0.333	0.352	-5.7	86	0.00
45 C	1,2-Dichloropropane	0.336	0.383	-14.0#	92	0.00
46 T	Dibromomethane	0.227	0.260	-14.5	91	0.00
47 T	Bromodichloromethane	0.451	0.527	-16.9	92	0.00
48 T	Methyl methacrylate	0.437	0.485	-11.0	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062322\
 Data File : VN073037.D
 Acq On : 23 Jun 2022 10:00
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 24 01:16:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.009	10.0	74	0.00
50 S	Toluene-d8	1.150	1.238	-7.7	85	0.00
51 T	4-Methyl-2-Pentanone	0.623	0.693	-11.2	87	0.00
52 CM	Toluene	0.807	0.892	-10.5#	88	0.00
53 T	t-1,3-Dichloropropene	0.485	0.542	-11.8	84	0.00
54 T	cis-1,3-Dichloropropene	0.530	0.578	-9.1	85	0.00
55 T	1,1,2-Trichloroethane	0.354	0.384	-8.5	91	0.00
56 T	Ethyl methacrylate	0.568	0.631	-11.1	86	0.00
57 T	1,3-Dichloropropane	0.588	0.650	-10.5	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.251	0.274	-9.2	80	0.00
59 T	2-Hexanone	0.482	0.525	-8.9	85	0.00
60 T	Dibromochloromethane	0.335	0.401	-19.7	89	0.00
61 T	1,2-Dibromoethane	0.349	0.393	-12.6	88	0.00
62 S	4-Bromofluorobenzene	0.414	0.456	-10.1	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.304	0.321	-5.6	88	0.00
65 PM	Chlorobenzene	0.976	1.066	-9.2	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.372	0.406	-9.1	88	0.00
67 C	Ethyl Benzene	1.719	1.922	-11.8#	89	0.00
68 T	m/p-Xylenes	0.674	0.732	-8.6	87	0.00
69 T	o-Xylene	0.674	0.746	-10.7	90	0.00
70 T	Styrene	1.082	1.236	-14.2	90	0.00
71 P	Bromoform	0.280	0.323	-15.4	87	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
73 T	Isopropylbenzene	3.868	4.126	-6.7	89	0.00
74 T	N-amyl acetate	2.007	2.062	-2.7	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.393	1.437	-3.2	88	0.00
76 T	1,2,3-Trichloropropane	1.212	1.263	-4.2	98	0.00
77 T	Bromobenzene	0.927	0.988	-6.6	89	0.00
78 T	n-propylbenzene	4.255	4.731	-11.2	90	0.00
79 T	2-Chlorotoluene	2.755	2.943	-6.8	91	0.00
80 T	1,3,5-Trimethylbenzene	3.177	3.458	-8.8	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.449	0.427	4.9	76	0.00
82 T	4-Chlorotoluene	2.607	2.873	-10.2	91	0.00
83 T	tert-Butylbenzene	2.910	3.027	-4.0	86	0.00
84 T	1,2,4-Trimethylbenzene	3.165	3.439	-8.7	90	0.00
85 T	sec-Butylbenzene	3.882	4.226	-8.9	89	0.00
86 T	p-Isopropyltoluene	3.143	3.444	-9.6	87	0.00
87 T	1,3-Dichlorobenzene	1.654	1.750	-5.8	90	0.00
88 T	1,4-Dichlorobenzene	1.674	1.791	-7.0	91	0.00
89 T	n-Butylbenzene	2.566	2.808	-9.4	86	0.00
90 T	Hexachloroethane	0.592	0.667	-12.7	91	0.00
91 T	1,2-Dichlorobenzene	1.725	1.827	-5.9	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.376	0.336	10.6	79	0.00
93 T	1,2,4-Trichlorobenzene	0.955	0.975	-2.1	85	0.00
94 T	Hexachlorobutadiene	0.416	0.394	5.3	79	0.00
95 T	Naphthalene	3.863	3.558	7.9	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062322\
Data File : VN073037.D
Acq On : 23 Jun 2022 10:00
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 24 01:16:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.020	0.958	6.1	82	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6