



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062322\
 Data File : VN073064.D
 Acq On : 23 Jun 2022 22:45
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N062322W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Fri Jun 24 04:54:59 2022



AutoFind: Scans 1832, 1833, 1834; Background Corrected with Scan 1825

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	12741	PASS
75	95	30	60	50.3	33616	PASS
95	95	100	100	100.0	66880	PASS
96	95	5	9	6.6	4427	PASS
173	174	0.00	2	1.0	496	PASS
174	95	50	100	74.4	49773	PASS
175	174	5	9	7.6	3790	PASS
176	174	95	101	99.4	49477	PASS
177	176	5	9	6.2	3066	PASS