

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062323\
 Data File : VN078377.D
 Acq On : 23 Jun 2023 14:52
 Operator : JC\MD
 Sample : 03379-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 230621026-03-TRIP-BLANK

Quant Time: Jun 24 01:20:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N062323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 24 00:59:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	54087	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	291454	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.870	117	229262	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.582	65	122050	29.966	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.900%
60) 4-Bromofluorobenzene	12.853	95	69439	21.498	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	71.667%
63) Toluene-d8	10.576	98	348563	31.379	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	104.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062323\
Data File : VN078377.D
Acq On : 23 Jun 2023 14:52
Operator : JC\MD
Sample : 03379-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
230621026-03-TRIP-BLANK

Quant Time: Jun 24 01:20:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N062323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 24 00:59:50 2023
Response via : Initial Calibration

