

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062323\
 Data File : VN078379.D
 Acq On : 23 Jun 2023 15:39
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 24 01:21:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N062323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 24 00:59:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	72540	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.112	114	376741	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	309426	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	166620	30.502	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.667%
60) 4-Bromofluorobenzene	12.853	95	98507	22.596	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	75.333%
63) Toluene-d8	10.571	98	458352	30.572	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062323\
Data File : VN078379.D
Acq On : 23 Jun 2023 15:39
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 24 01:21:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N062323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 24 00:59:50 2023
Response via : Initial Calibration

