

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062323\
 Data File : VN078383.D
 Acq On : 23 Jun 2023 17:15
 Operator : JC\MD
 Sample : 03395-01DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 001-WILLETS-PT-BLVD(JUNE)DL

Quant Time: Jun 24 01:23:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N062323W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 24 00:59:50 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.824	128	92668	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.112	114	505038	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.870	117	413130	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.582	65	205892	29.505	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.333%
60) 4-Bromofluorobenzene	12.853	95	142326	24.452	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	81.500%
63) Toluene-d8	10.576	98	605740	30.261	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	100.867%
Target Compounds						
					Qvalue	
15) Acetone	4.447	58	37651	60.151	ug/l	96
30) 2-Butanone	7.500	43	15232	5.551	ug/l #	77
58) 4-Methyl-2-Pentanone	10.447	43	3684	0.668	ug/l #	77
62) Toluene	10.635	91	590062	24.541	ug/l	99
82) p-Isopropyltoluene	13.729	119	4464	0.296	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062323\
Data File : VN078383.D
Acq On : 23 Jun 2023 17:15
Operator : JC\MD
Sample : 03395-01DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
001-WILLETS-PT-BLVD(JUNE)DL

Quant Time: Jun 24 01:23:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N062323W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 24 00:59:50 2023
Response via : Initial Calibration

