

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062419\
 Data File : VN056453.D
 Acq On : 24 Jun 2019 14:12
 Operator : JC/SP
 Sample : PB120834ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB120834ZHE#08

Quant Time: Jun 25 01:54:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	375556	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	638809	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	639193	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	187771	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	212876	51.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.28%	
35) Dibromofluoromethane	7.59	113	184466	46.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.58%	
50) Toluene-d8	10.09	98	794721	51.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.50%	
62) 4-Bromofluorobenzene	12.40	95	256941	49.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.42%	
Target Compounds						
43) Isopropyl Acetate	8.17	43	15996	1.877	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062419\
Data File : VN056453.D
Acq On : 24 Jun 2019 14:12
Operator : JC/SP
Sample : PB120834ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB120834ZHE#08

Quant Time: Jun 25 01:54:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

