

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062419\
 Data File : VN056460.D
 Acq On : 24 Jun 2019 16:56
 Operator : JC/SP
 Sample : K3482-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Jun 25 02:22:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	377397	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	639819	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	650995	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	200973	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	214034	51.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.34%	
35) Dibromofluoromethane	7.59	113	183695	46.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.04%	
50) Toluene-d8	10.09	98	808625	52.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.14%	
62) 4-Bromofluorobenzene	12.40	95	272522	52.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.28%	
Target Compounds						
20) Methylene Chloride	4.55	84	103378	25.736	ug/l	99
27) cis-1,2-Dichloroethene	6.83	96	7060	1.679	ug/l	97
43) Isopropyl Acetate	8.17	43	18219	2.135	ug/l #	85
44) Trichloroethene	8.83	130	11830	2.535	ug/l	98
95) Naphthalene	15.14	128	5409	5.771	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062419\
Data File : VN056460.D
Acq On : 24 Jun 2019 16:56
Operator : JC/SP
Sample : K3482-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Jun 25 02:22:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

