

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062419\
 Data File : VN056441.D
 Acq On : 24 Jun 2019 9:21
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 25 01:30:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	110	0.00
2 T	Dichlorodifluoromethane	0.423	0.422	0.2	108	0.00
3 P	Chloromethane	0.576	0.513	10.9	101	0.00
4 C	Vinyl Chloride	0.557	0.529	5.0#	106	0.00
5 T	Bromomethane	0.326	0.292	10.4	94	0.00
6 T	Chloroethane	0.322	0.306	5.0	105	0.00
7 T	Trichlorofluoromethane	0.720	0.715	0.7	112	0.00
8 T	Diethyl Ether	0.305	0.294	3.6	110	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.459	0.459	0.0	114	0.00
10 T	Methyl Iodide	0.621	0.611	1.6	104	0.00
11 T	Tert butyl alcohol	0.093	0.089	4.3	106	0.00
12 CM	1,1-Dichloroethene	0.442	0.439	0.7#	110	0.00
13 T	Acrolein	0.079	0.078	1.3	111	0.00
14 T	Allyl chloride	0.735	0.741	-0.8	110	0.00
15 T	Acrylonitrile	0.258	0.246	4.7	105	0.00
16 T	Acetone	0.278	0.294	-5.8	119	0.00
17 T	Carbon Disulfide	1.005	0.976	2.9	107	0.00
18 T	Methyl Acetate	0.696	0.522	25.0	109	0.00
19 T	Methyl tert-butyl Ether	1.462	1.458	0.3	112	0.00
20 T	Methylene Chloride	0.571	0.505	11.6	106	0.00
21 T	trans-1,2-Dichloroethene	0.476	0.466	2.1	109	0.00
22 T	Diisopropyl ether	1.544	1.549	-0.3	112	0.00
23 T	Vinyl Acetate	1.218	1.285	-5.5	110	0.00
24 P	1,1-Dichloroethane	0.886	0.884	0.2	111	0.00
25 T	2-Butanone	0.367	0.368	-0.3	109	0.00
26 T	2,2-Dichloropropane	0.616	0.672	-9.1	121	0.00
27 T	cis-1,2-Dichloroethene	0.557	0.558	-0.2	110	0.00
28 T	Bromochloromethane	0.438	0.438	0.0	110	0.00
29 T	Tetrahydrofuran	0.225	0.214	4.9	105	0.00
30 C	Chloroform	0.875	0.872	0.3#	112	0.00
31 T	Cyclohexane	0.856	0.794	7.2	108	0.00
32 T	1,1,1-Trichloroethane	0.693	0.721	-4.0	113	0.00
33 S	1,2-Dichloroethane-d4	0.554	0.558	-0.7	110	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	107	0.00
35 S	Dibromofluoromethane	0.312	0.330	-5.8	112	0.00
36 T	1,1-Dichloropropene	0.440	0.434	1.4	109	0.00
37 T	Ethyl Acetate	0.432	0.434	-0.5	103	0.00
38 T	Carbon Tetrachloride	0.385	0.405	-5.2	115	0.00
39 T	Methylcyclohexane	0.543	0.546	-0.6	111	0.00
40 TM	Benzene	1.354	1.342	0.9	109	0.00
41 T	Methacrylonitrile	0.191	0.181	5.2	107	0.00
42 TM	1,2-Dichloroethane	0.447	0.436	2.5	109	0.00
43 T	Isopropyl Acetate	0.667	0.690	-3.4	109	0.00
44 TM	Trichloroethene	0.365	0.369	-1.1	110	0.00
45 C	1,2-Dichloropropane	0.362	0.365	-0.8#	110	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062419\
 Data File : VN056441.D
 Acq On : 24 Jun 2019 9:21
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 25 01:30:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 03:45:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.227	0.232	-2.2	110	0.00
47 T	Bromodichloromethane	0.402	0.442	-10.0	116	0.00
48 T	Methyl methacrylate	0.334	0.330	1.2	108	0.00
49 T	1,4-Dioxane	0.007	0.007	0.0	108	0.00
50 S	Toluene-d8	1.214	1.235	-1.7	108	0.00
51 T	4-Methyl-2-Pentanone	0.430	0.423	1.6	104	0.00
52 CM	Toluene	0.828	0.838	-1.2#	111	0.00
53 T	t-1,3-Dichloropropene	0.405	0.473	-16.8	116	0.00
54 T	cis-1,3-Dichloropropene	0.484	0.543	-12.2	113	0.00
55 T	1,1,2-Trichloroethane	0.333	0.340	-2.1	110	0.00
56 T	Ethyl methacrylate	0.486	0.510	-4.9	109	0.00
57 T	1,3-Dichloropropane	0.559	0.567	-1.4	110	0.00
58 T	2-Chloroethyl Vinyl ether	0.175	0.174	0.6	105	0.00
59 T	2-Hexanone	0.316	0.309	2.2	104	0.00
60 T	Dibromochloromethane	0.302	0.348	-15.2	116	0.00
61 T	1,2-Dibromoethane	0.340	0.350	-2.9	109	0.00
62 S	4-Bromofluorobenzene	0.405	0.430	-6.2	111	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.396	0.392	1.0	111	0.00
65 PM	Chlorobenzene	1.005	1.011	-0.6	110	0.00
66 T	1,1,1,2-Tetrachloroethane	0.340	0.374	-10.0	115	0.00
67 C	Ethyl Benzene	1.764	1.781	-1.0#	110	0.00
68 T	m/p-Xylenes	0.646	0.670	-3.7	110	0.00
69 T	o-Xylene	0.640	0.665	-3.9	112	0.00
70 T	Styrene	1.022	1.101	-7.7	110	0.00
71 P	Bromoform	0.236	0.275	-16.5	117	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
73 T	Isopropylbenzene	4.403	4.095	7.0	112	0.00
74 T	N-amyl acetate	1.591	1.489	6.4	108	0.00
75 P	1,1,2,2-Tetrachloroethane	1.374	1.218	11.4	107	0.00
76 T	1,2,3-Trichloropropane	1.037	0.959	7.5	106	0.00
77 T	Bromobenzene	1.118	1.052	5.9	109	0.00
78 T	n-propylbenzene	4.579	4.482	2.1	112	0.00
79 T	2-Chlorotoluene	2.893	2.710	6.3	110	0.00
80 T	1,3,5-Trimethylbenzene	3.589	3.381	5.8	111	0.00
81 T	trans-1,4-Dichloro-2-butene	0.292	0.321	-9.9	121	0.00
82 T	4-Chlorotoluene	2.704	2.600	3.8	109	0.00
83 T	tert-Butylbenzene	3.279	2.987	8.9	112	0.00
84 T	1,2,4-Trimethylbenzene	3.366	3.248	3.5	109	0.00
85 T	sec-Butylbenzene	4.024	3.891	3.3	112	0.00
86 T	p-Isopropyltoluene	3.454	3.462	-0.2	112	0.00
87 T	1,3-Dichlorobenzene	1.694	1.723	-1.7	111	0.00
88 T	1,4-Dichlorobenzene	1.641	1.627	0.9	110	0.00
89 T	n-Butylbenzene	2.615	2.673	-2.2	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062419\
Data File : VN056441.D
Acq On : 24 Jun 2019 9:21
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 25 01:30:28 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061919W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 03:45:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.538	0.571	-6.1	120	0.00
91 T	1,2-Dichlorobenzene	1.681	1.667	0.8	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.185	0.188	-1.6	106	0.00
93 T	1,2,4-Trichlorobenzene	0.703	0.751	-6.8	105	0.00
94 T	Hexachlorobutadiene	0.644	0.615	4.5	113	0.00
95 T	Naphthalene	1.854	2.008	-8.3	106	0.00
96 T	1,2,3-Trichlorobenzene	0.762	0.826	-8.4	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6