

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062420\
 Data File : VN062088.D
 Acq On : 25 Jun 2020 1:51
 Operator : JC/MD
 Sample : PB129725ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129725ZHE#01

Quant Time: Jun 25 05:53:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	229069	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	435656	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	443619	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	165132	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	164738	51.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.50%	
35) Dibromofluoromethane	7.55	113	138573	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
50) Toluene-d8	10.06	98	568056	50.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.68%	
62) 4-Bromofluorobenzene	12.38	95	188865	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
Target Compounds						
16) Acetone	3.78	43	13782	9.766	ug/l	99
18) Methyl Acetate	4.28	43	2868	1.122	ug/l #	59
20) Methylene Chloride	4.51	84	9336	2.960	ug/l	95
43) Isopropyl Acetate	8.13	43	59998	10.150	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062420\
Data File : VN062088.D
Acq On : 25 Jun 2020 1:51
Operator : JC/MD
Sample : PB129725ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129725ZHE#01

Quant Time: Jun 25 05:53:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

