

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062420\
 Data File : VN062090.D
 Acq On : 25 Jun 2020 2:39
 Operator : JC/MD
 Sample : PB129725ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129725ZHE#03

Quant Time: Jun 25 05:56:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	221802	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	433028	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	422376	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	149038	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	161406	51.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.72%	
35) Dibromofluoromethane	7.55	113	136322	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
50) Toluene-d8	10.06	98	556127	50.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.16%	
62) 4-Bromofluorobenzene	12.38	95	179456	47.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.66%	
Target Compounds						
16) Acetone	3.78	43	14196	10.652	ug/l	93
18) Methyl Acetate	4.29	43	4041	1.633	ug/l	90
20) Methylene Chloride	4.50	84	13951	4.686	ug/l	96
43) Isopropyl Acetate	8.13	43	57330	9.758	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062420\
Data File : VN062090.D
Acq On : 25 Jun 2020 2:39
Operator : JC/MD
Sample : PB129725ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129725ZHE#03

Quant Time: Jun 25 05:56:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

