

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062420\
 Data File : VN062091.D
 Acq On : 25 Jun 2020 3:03
 Operator : JC/MD
 Sample : PB129725ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB129725ZHE#04

Quant Time: Jun 25 05:58:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	210352	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	414851	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	402500	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	147890	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	152736	51.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.48%	
35) Dibromofluoromethane	7.55	113	129411	48.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.24%	
50) Toluene-d8	10.06	98	529697	49.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.58%	
62) 4-Bromofluorobenzene	12.38	95	170918	47.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.10%	
Target Compounds						
16) Acetone	3.78	43	13490	10.681	ug/l	95
18) Methyl Acetate	4.28	43	4491	1.914	ug/l #	67
20) Methylene Chloride	4.50	84	13792	4.894	ug/l	97
43) Isopropyl Acetate	8.13	43	57782	10.266	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062420\
Data File : VN062091.D
Acq On : 25 Jun 2020 3:03
Operator : JC/MD
Sample : PB129725ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB129725ZHE#04

Quant Time: Jun 25 05:58:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

