

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062420\
 Data File : VN062096.D
 Acq On : 25 Jun 2020 5:04
 Operator : JC/MD
 Sample : L2818-10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-05-062320

Quant Time: Jun 25 06:13:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	208232	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	406897	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	417315	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	154006	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	154409	52.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.68%	
35) Dibromofluoromethane	7.55	113	131959	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
50) Toluene-d8	10.06	98	536552	51.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.84%	
62) 4-Bromofluorobenzene	12.38	95	175389	49.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.50%	
Target Compounds						
16) Acetone	3.78	43	21934	20.189	ug/l	91
18) Methyl Acetate	4.28	43	4735	2.038	ug/l #	67
20) Methylene Chloride	4.50	84	35283	12.992	ug/l	98
43) Isopropyl Acetate	8.13	43	54510	9.874	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062420\
Data File : VN062096.D
Acq On : 25 Jun 2020 5:04
Operator : JC/MD
Sample : L2818-10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-05-062320

Quant Time: Jun 25 06:13:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

