

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062420\
 Data File : VN062077.D
 Acq On : 24 Jun 2020 21:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 25 05:27:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.565	0.518	8.3	93	0.00
3 P	Chloromethane	0.744	0.648	12.9	92	0.00
4 C	Vinyl Chloride	0.890	0.815	8.4#	94	0.00
5 T	Bromomethane	0.575	0.315	45.2#	56	0.00
6 T	Chloroethane	0.608	0.573	5.8	95	0.00
7 T	Trichlorofluoromethane	1.097	1.009	8.0	104	0.00
8 T	Diethyl Ether	0.383	0.369	3.7	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.574	0.525	8.5	95	0.00
10 T	Methyl Iodide	0.647	0.369	43.0#	55	0.00
11 T	Tert butyl alcohol	0.101	0.095	5.9	95	0.00
12 CM	1,1-Dichloroethene	0.574	0.545	5.1#	98	0.00
13 T	Acrolein	0.099	0.089	10.1	88	0.00
14 T	Allyl chloride	0.782	0.751	4.0	95	0.00
15 T	Acrylonitrile	0.275	0.266	3.3	96	0.00
16 T	Acetone	0.251	0.210	16.3	90	0.00
17 T	Carbon Disulfide	1.833	1.604	12.5	95	0.00
18 T	Methyl Acetate	0.558	0.528	5.4	101	0.00
19 T	Methyl tert-butyl Ether	1.835	1.821	0.8	98	0.00
20 T	Methylene Chloride	0.708	0.645	8.9	97	0.00
21 T	trans-1,2-Dichloroethene	0.654	0.602	8.0	95	0.00
22 T	Diisopropyl ether	1.780	1.706	4.2	97	0.00
23 T	Vinyl Acetate	1.453	1.457	-0.3	97	0.00
24 P	1,1-Dichloroethane	1.146	1.066	7.0	96	0.00
25 T	2-Butanone	0.343	0.333	2.9	96	0.00
26 T	2,2-Dichloropropane	0.955	0.834	12.7	88	0.00
27 T	cis-1,2-Dichloroethene	0.736	0.698	5.2	98	0.00
28 T	Bromochloromethane	0.497	0.478	3.8	99	0.00
29 T	Tetrahydrofuran	0.225	0.225	0.0	97	0.00
30 C	Chloroform	1.166	1.093	6.3#	96	0.00
31 T	Cyclohexane	0.969	0.909	6.2	97	0.00
32 T	1,1,1-Trichloroethane	0.983	0.941	4.3	97	0.00
33 S	1,2-Dichloroethane-d4	0.702	0.675	3.8	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.321	0.305	5.0	100	0.00
36 T	1,1-Dichloropropene	0.503	0.464	7.8	97	0.00
37 T	Ethyl Acetate	0.416	0.402	3.4	98	0.00
38 T	Carbon Tetrachloride	0.474	0.441	7.0	97	0.00
39 T	Methylcyclohexane	0.551	0.527	4.4	98	0.00
40 TM	Benzene	1.532	1.422	7.2	96	0.00
41 T	Methacrylonitrile	0.193	0.191	1.0	106	0.00
42 TM	1,2-Dichloroethane	0.518	0.476	8.1	95	0.00
43 T	Isopropyl Acetate	0.678	0.666	1.8	100	0.00
44 TM	Trichloroethene	0.386	0.353	8.5	97	0.00
45 C	1,2-Dichloropropane	0.378	0.359	5.0#	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062420\
 Data File : VN062077.D
 Acq On : 24 Jun 2020 21:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 25 05:27:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 24 15:33:00 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.256	0.242	5.5	97	0.00
47 T	Bromodichloromethane	0.517	0.491	5.0	95	0.00
48 T	Methyl methacrylate	0.304	0.312	-2.6	102	0.00
49 T	1,4-Dioxane	0.007	0.007	0.0	91	0.00
50 S	Toluene-d8	1.282	1.222	4.7	100	0.00
51 T	4-Methyl-2-Pentanone	0.396	0.404	-2.0	99	0.00
52 CM	Toluene	0.935	0.913	2.4#	96	0.00
53 T	t-1,3-Dichloropropene	0.569	0.542	4.7	95	0.00
54 T	cis-1,3-Dichloropropene	0.606	0.587	3.1	96	0.00
55 T	1,1,2-Trichloroethane	0.367	0.353	3.8	95	0.00
56 T	Ethyl methacrylate	0.490	0.517	-5.5	99	0.00
57 T	1,3-Dichloropropane	0.632	0.606	4.1	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.184	0.201	-9.2	104	0.00
59 T	2-Hexanone	0.286	0.297	-3.8	98	0.00
60 T	Dibromochloromethane	0.389	0.377	3.1	97	0.00
61 T	1,2-Dibromoethane	0.378	0.364	3.7	97	0.00
62 S	4-Bromofluorobenzene	0.433	0.419	3.2	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.342	0.313	8.5	96	0.00
65 PM	Chlorobenzene	1.098	1.022	6.9	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.380	0.359	5.5	94	0.00
67 C	Ethyl Benzene	1.854	1.846	0.4#	96	0.00
68 T	m/p-Xylenes	0.717	0.706	1.5	94	0.00
69 T	o-Xylene	0.677	0.678	-0.1	95	0.00
70 T	Styrene	1.129	1.157	-2.5	96	0.00
71 P	Bromoform	0.270	0.267	1.1	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.870	3.687	4.7	96	0.00
74 T	N-amyl acetate	1.265	1.341	-6.0	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.288	1.178	8.5	97	0.00
76 T	1,2,3-Trichloropropane	1.177	0.972	17.4	87	0.00
77 T	Bromobenzene	0.961	0.875	8.9	95	0.00
78 T	n-propylbenzene	4.427	4.253	3.9	95	0.00
79 T	2-Chlorotoluene	2.676	2.511	6.2	96	0.00
80 T	1,3,5-Trimethylbenzene	3.138	3.068	2.2	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.403	0.338	16.1	83	0.00
82 T	4-Chlorotoluene	2.768	2.583	6.7	95	0.00
83 T	tert-Butylbenzene	2.685	2.573	4.2	96	0.00
84 T	1,2,4-Trimethylbenzene	3.134	3.063	2.3	96	0.00
85 T	sec-Butylbenzene	3.548	3.434	3.2	96	0.00
86 T	p-Isopropyltoluene	3.109	3.076	1.1	96	0.00
87 T	1,3-Dichlorobenzene	1.687	1.569	7.0	95	0.00
88 T	1,4-Dichlorobenzene	1.739	1.576	9.4	95	0.00
89 T	n-Butylbenzene	2.584	2.547	1.4	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062420\
Data File : VN062077.D
Acq On : 24 Jun 2020 21:02
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jun 25 05:27:44 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 24 15:33:00 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.610	0.531	13.0	90	0.00
91 T	1,2-Dichlorobenzene	1.636	1.539	5.9	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.206	0.201	2.4	95	0.00
93 T	1,2,4-Trichlorobenzene	0.715	0.732	-2.4	99	0.00
94 T	Hexachlorobutadiene	0.310	0.287	7.4	97	0.00
95 T	Naphthalene	2.084	2.209	-6.0	100	0.00
96 T	1,2,3-Trichlorobenzene	0.731	0.733	-0.3	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6