


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN062420\
Data File : VN062078.D
Acq On : 24 Jun 2020 21:50
Operator : JC/MD
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N062420W.M
Title : SW846 8260
Last Update : Wed Jun 24 15:33:00 2020



AutoFind: Scans 3297, 3298, 3299; Background Corrected with Scan 3283

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.2	11784	PASS
75	95	30	60	49.6	32170	PASS
95	95	100	100	100.0	64893	PASS
96	95	5	9	6.8	4407	PASS
173	174	0.00	2	1.5	727	PASS
174	95	50	100	75.2	48784	PASS
175	174	5	9	8.4	4078	PASS
176	174	95	101	95.6	46661	PASS
177	176	5	9	6.2	2882	PASS