

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062421\
 Data File : VN067479.D
 Acq On : 24 Jun 2021 13:22
 Operator : JC/MD
 Sample : VN0624MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0624MBL01

Quant Time: Jun 25 02:13:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 24 12:18:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	270588	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	574769	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	503475	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	157448	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	261361	47.196	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.400%
35) Dibromofluoromethane	8.024	113	175965	49.693	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.380%
50) Toluene-d8	10.443	98	754438	49.954	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.900%
62) 4-Bromofluorobenzene	12.734	95	249090	46.292	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062421\
Data File : VN067479.D
Acq On : 24 Jun 2021 13:22
Operator : JC/MD
Sample : VN0624MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0624MBL01

Quant Time: Jun 25 02:13:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 24 12:18:45 2021
Response via : Initial Calibration

