

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN062421\  
 Data File : VN067486.D  
 Acq On : 24 Jun 2021 16:30  
 Operator : JC/MD  
 Sample : M2804-01 10X  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 17 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 TT101D2-20210621

Quant Time: Jun 25 02:14:57 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N062421W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Jun 24 12:18:45 2021  
 Response via : Initial Calibration

| Compound                     | R.T.           | QIon | Response   | Conc     | Units  | Dev(Min) |
|------------------------------|----------------|------|------------|----------|--------|----------|
| -----                        |                |      |            |          |        |          |
| Internal Standards           |                |      |            |          |        |          |
| 1) Pentafluorobenzene        | 8.086          | 168  | 251662     | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 8.971          | 114  | 534522     | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 11.749         | 117  | 475849     | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 13.678         | 152  | 155577     | 50.000   | ug/l   | 0.00     |
|                              |                |      |            |          |        |          |
| System Monitoring Compounds  |                |      |            |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 8.440          | 65   | 249363     | 48.416   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 61 - 141 |      | Recovery = | 96.840%  |        |          |
| 35) Dibromofluoromethane     | 8.024          | 113  | 161003     | 48.892   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 69 - 133 |      | Recovery = | 97.780%  |        |          |
| 50) Toluene-d8               | 10.443         | 98   | 712962     | 50.762   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 65 - 126 |      | Recovery = | 101.520% |        |          |
| 62) 4-Bromofluorobenzene     | 12.733         | 95   | 236962     | 47.354   | ug/l   | 0.00     |
| Spiked Amount 50.000         | Range 58 - 135 |      | Recovery = | 94.700%  |        |          |
|                              |                |      |            |          |        |          |
| Target Compounds             |                |      |            |          |        |          |
|                              |                |      |            |          | Qvalue |          |
| 9) 1,1,2-Trichlorotrifluo... | 4.208          | 101  | 11504      | 3.301    | ug/l # | 52       |
| 12) 1,1-Dichloroethene       | 4.181          | 96   | 3485       | 1.011    | ug/l   | 83       |
| 44) Trichloroethene          | 9.217          | 130  | 482566     | 138.116  | ug/l   | 93       |
| 64) Tetrachloroethene        | 10.977         | 164  | 1304       | 0.421    | ug/l   | 89       |
| 89) n-Butylbenzene           | 13.946         | 91   | 5787       | 0.508    | ug/l   | 97       |
| 95) Naphthalene              | 15.509         | 128  | 15494      | 5.002    | ug/l   | 99       |
| -----                        |                |      |            |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN062421\  
Data File : VN067486.D  
Acq On : 24 Jun 2021 16:30  
Operator : JC/MD  
Sample : M2804-01 10X  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 17 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
TT101D2-20210621

Quant Time: Jun 25 02:14:57 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N062421W.M  
Quant Title : SW846 8260  
QLast Update : Thu Jun 24 12:18:45 2021  
Response via : Initial Calibration

