

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062421\
 Data File : VN067491.D
 Acq On : 24 Jun 2021 18:32
 Operator : JC/MD
 Sample : M2804-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP07-20210621

Manual Integrations
 APPROVED

MMDadoda
 6/25/2021 6:23:02 PM

Quant Time: Jun 25 02:15:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 24 12:18:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	239935	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	510141	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	459806	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	152576	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	244931	49.879	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	99.760%		
35) Dibromofluoromethane	8.024	113	156000	49.636	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.280%		
50) Toluene-d8	10.443	98	681101	50.811	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	101.620%		
62) 4-Bromofluorobenzene	12.734	95	231001	48.369	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	96.740%		
Target Compounds					Qvalue	
16) Acetone	4.302	43	5362m	2.853	ug/l	
44) Trichloroethene	9.220	130	78683	23.596	ug/l	93
64) Tetrachloroethene	10.982	164	5128	1.715	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062421\
Data File : VN067491.D
Acq On : 24 Jun 2021 18:32
Operator : JC/MD
Sample : M2804-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP07-20210621

Manual Integrations
APPROVED

MMDadoda
6/25/2021 6:23:02 PM

Quant Time: Jun 25 02:15:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 24 12:18:45 2021
Response via : Initial Calibration

