

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062421\
 Data File : VN067494.D
 Acq On : 24 Jun 2021 19:44
 Operator : JC/MD
 Sample : M2828-11
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ERM-19

Manual Integrations
 APPROVED

MMDadoda
 6/25/2021 6:23:03 PM

Quant Time: Jun 25 02:16:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 24 12:18:45 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	231501	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	495848	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	445201	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	159040	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	236387	49.893	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	99.780%		
35) Dibromofluoromethane	8.024	113	151488	49.590	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.180%		
50) Toluene-d8	10.443	98	667266	51.214	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	102.420%		
62) 4-Bromofluorobenzene	12.734	95	228566	49.238	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	98.480%		
Target Compounds						
16) Acetone	4.293	43	5631m	3.105	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062421\
Data File : VN067494.D
Acq On : 24 Jun 2021 19:44
Operator : JC/MD
Sample : M2828-11
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ERM-19

Manual Integrations
APPROVED

MMDadoda
6/25/2021 6:23:03 PM

Quant Time: Jun 25 02:16:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062421W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 24 12:18:45 2021
Response via : Initial Calibration

