


t-1,3-Dichloropropene

Response Ratio




$$\text{Response} = 7.372\text{e-}001 * \text{Amt} - 2.954\text{e-}002$$

Coef of Det (r^2) = 0.999161 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062421W.M

Calibration Table Last Updated: Thu Jun 24 12:18:45 2021


Bromoform



Response = 2.855e-001 * Amt - 1.517e-002
 Coef of Det (r^2) = 0.998598 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062421W.M
 Calibration Table Last Updated: Thu Jun 24 12:18:45 2021

1,2,4-Trichlorobenzene

Response Ratio



$$\text{Response} = 8.815\text{e-}001 * \text{Amt} - 4.106\text{e-}002$$


Coef of Det (r^2) = 0.999202 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062421W.M

Calibration Table Last Updated: Thu Jun 24 12:18:45 2021

Naphthalene

Response Ratio



$$\text{Response} = 2.990\text{e}+000 * \text{Amt} - 1.995\text{e}-001$$


Coef of Det (r^2) = 0.998436 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062421W.M

Calibration Table Last Updated: Thu Jun 24 12:18:45 2021

1,2,3-Trichlorobenzene

Response Ratio



$$\text{Response} = 8.430\text{e-}001 * \text{Amt} - 4.259\text{e-}002$$


Coef of Det (r^2) = 0.999049 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062421W.M

Calibration Table Last Updated: Thu Jun 24 12:18:45 2021

2-Chloroethyl Vinyl ether


Response Ratio



Response = 1.881e-001 * Amt - 6.480e-002
 Coef of Det (r^2) = 0.997298 Curve Fit: Linear
 Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062421W.M
 Calibration Table Last Updated: Thu Jun 24 12:18:45 2021

Dibromochloromethane

Response Ratio



$$\text{Response} = 3.951\text{e-}001 * \text{Amt} - 1.605\text{e-}002$$


Coef of Det (r^2) = 0.998900 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062421W.M

Calibration Table Last Updated: Thu Jun 24 12:18:45 2021

1,1,1,2-Tetrachloroethane

Response Ratio



$$\text{Response} = 3.792\text{e-}001 * \text{Amt} - 5.412\text{e-}003$$

Coef of Det (r^2) = 0.999210 Curve Fit: Linear

Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N062421W.M

Calibration Table Last Updated: Thu Jun 24 12:18:45 2021