

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062422\
 Data File : VN073090.D
 Acq On : 24 Jun 2022 18:11
 Operator : JC\MD
 Sample : N3455-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-EXC-2-2-GR

Quant Time: Jun 24 23:40:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	292266	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	468739	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	438373	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	213179	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	178090	45.071	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.140%
35) Dibromofluoromethane	7.957	113	151126	52.162	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.320%
50) Toluene-d8	10.374	98	498747	46.263	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.520%
62) 4-Bromofluorobenzene	12.668	95	213820	55.157	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.320%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	66419	30.579	ug/l	95
20) Methylene Chloride	5.016	84	146880	43.566	ug/l	90
43) Isopropyl Acetate	8.498	43	28605	3.191	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062422\
Data File : VN073090.D
Acq On : 24 Jun 2022 18:11
Operator : JC\MD
Sample : N3455-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-EXC-2-2-GR

Quant Time: Jun 24 23:40:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

