

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062422\
 Data File : VN073095.D
 Acq On : 24 Jun 2022 20:14
 Operator : JC\MD
 Sample : N3455-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP-EXC-2-1-GR

Quant Time: Jun 24 23:41:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 18 06:22:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	276054	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	457624	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	429277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	208387	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.375	65	173555	46.502	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.000%
35) Dibromofluoromethane	7.951	113	145473	51.431	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.860%
50) Toluene-d8	10.380	98	484311	46.015	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.040%
62) 4-Bromofluorobenzene	12.668	95	206081	54.452	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.900%
Target Compounds						
					Qvalue	
16) Acetone	4.228	43	58703	28.614	ug/l	89
18) Methyl Acetate	4.781	43	5580	1.021	ug/l #	79
20) Methylene Chloride	5.022	84	145099	45.576	ug/l	92
43) Isopropyl Acetate	8.498	43	26035	2.975	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062422\
Data File : VN073095.D
Acq On : 24 Jun 2022 20:14
Operator : JC\MD
Sample : N3455-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP-EXC-2-1-GR

Quant Time: Jun 24 23:41:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061722W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 18 06:22:45 2022
Response via : Initial Calibration

