

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062423\
 Data File : VN078394.D
 Acq On : 24 Jun 2023 15:01
 Operator : JC\MD
 Sample : 03393-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SF-10

Quant Time: Jun 26 01:50:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 07 05:05:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	437044	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	810745	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	733366	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	220548	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	298513	51.228	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.460%		
35) Dibromofluoromethane	8.177	113	284675	55.533	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 111.060%		
50) Toluene-d8	10.571	98	953495	48.249	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 96.500%		
62) 4-Bromofluorobenzene	12.859	95	297174	48.372	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 96.740%		
Target Compounds						
						Qvalue
16) Acetone	4.442	43	82680	49.295	ug/l	98
20) Methylene Chloride	5.294	84	31005	5.594	ug/l	96
25) 2-Butanone	7.500	43	19482	7.553	ug/l #	87
37) Ethyl Acetate	7.577	43	44400	7.730	ug/l	96
43) Isopropyl Acetate	8.700	43	236533	21.532	ug/l #	81
51) 4-Methyl-2-Pentanone	10.447	43	146308	25.292	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062423\
Data File : VN078394.D
Acq On : 24 Jun 2023 15:01
Operator : JC\MD
Sample : 03393-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SF-10

Quant Time: Jun 26 01:50:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060623W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 07 05:05:00 2023
Response via : Initial Calibration

