

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062424\
 Data File : VN082693.D
 Acq On : 24 Jun 2024 10:57
 Operator : JC\MD
 Sample : VN0624WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0624WBL01

Quant Time: Jun 25 02:24:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	155114	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	292212	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	280580	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	117095	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	128699	51.739	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.480%
35) Dibromofluoromethane	8.165	113	96400	48.629	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.260%
50) Toluene-d8	10.571	98	345167	46.044	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.080%
62) 4-Bromofluorobenzene	12.853	95	110335	42.691	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	85.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062424\
Data File : VN082693.D
Acq On : 24 Jun 2024 10:57
Operator : JC\MD
Sample : VN0624WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0624WBL01

Quant Time: Jun 25 02:24:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

