

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062424\
 Data File : VN082697.D
 Acq On : 24 Jun 2024 13:19
 Operator : JC\MD
 Sample : P3034-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EB-01-062124

Quant Time: Jun 25 02:27:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153736	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	282571	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	271410	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	111724	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	127479	51.708	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.420%
35) Dibromofluoromethane	8.171	113	96289	50.231	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.460%
50) Toluene-d8	10.571	98	345366	47.642	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.280%
62) 4-Bromofluorobenzene	12.853	95	109820	43.941	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	87.880%
Target Compounds						
16) Acetone	4.436	43	11651	10.391	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062424\
Data File : VN082697.D
Acq On : 24 Jun 2024 13:19
Operator : JC\MD
Sample : P3034-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EB-01-062124

Quant Time: Jun 25 02:27:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

