

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062424\
 Data File : VN082701.D
 Acq On : 24 Jun 2024 14:56
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jun 25 02:29:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 18 04:52:44 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	140232	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	253082	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	246687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	102009	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	122501	54.474	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.940%
35) Dibromofluoromethane	8.171	113	90950	52.974	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.940%
50) Toluene-d8	10.571	98	332324	51.185	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.360%
62) 4-Bromofluorobenzene	12.847	95	104806	46.821	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	93.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062424\
Data File : VN082701.D
Acq On : 24 Jun 2024 14:56
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jun 25 02:29:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N061724W.M
Quant Title : SW846 8260
QLast Update : Tue Jun 18 04:52:44 2024
Response via : Initial Calibration

