

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062425\
 Data File : VN087150.D
 Acq On : 24 Jun 2025 12:44
 Operator : JC\MD
 Sample : VN0624WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0624WBL01

Quant Time: Jun 25 04:14:24 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	149121	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	287883	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	275522	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	132163	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	112496	56.345	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 112.680%	
35) Dibromofluoromethane	8.171	113	90627	53.121	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.240%	
50) Toluene-d8	10.565	98	357644	52.954	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.900%	
62) 4-Bromofluorobenzene	12.847	95	131459	52.390	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.780%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062425\
Data File : VN087150.D
Acq On : 24 Jun 2025 12:44
Operator : JC\MD
Sample : VN0624WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0624WBL01

Quant Time: Jun 25 04:14:24 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

