

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062425\
 Data File : VN087154.D
 Acq On : 24 Jun 2025 14:23
 Operator : JC\MD
 Sample : Q2366-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 250618063-05-TRIP-BLANK

Quant Time: Jun 25 04:18:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	142734	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	293524	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	271099	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	118196	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	115605	60.493	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 120.980%	
35) Dibromofluoromethane	8.177	113	90970	52.297	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.600%	
50) Toluene-d8	10.571	98	366978	53.292	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.580%	
62) 4-Bromofluorobenzene	12.847	95	126571	49.472	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 98.940%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN062425\
Data File : VN087154.D
Acq On : 24 Jun 2025 14:23
Operator : JC\MD
Sample : Q2366-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
250618063-05-TRIP-BLANK

Quant Time: Jun 25 04:18:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

