

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062518\
 Data File : VN049517.D
 Acq On : 25 Jun 2018 18:04
 Operator : MD\SY
 Sample : PB110537ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110537ZHE#02

Quant Time: Jun 26 02:11:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1214677	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1886533	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1586124	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	417139	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	804104	49.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.82%	
35) Dibromofluoromethane	7.59	113	716610	46.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.90%	
50) Toluene-d8	10.09	98	2624632	45.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.32%	
62) 4-Bromofluorobenzene	12.40	95	640212	33.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	66.64%	
Target Compounds						
16) Acetone	3.82	43	58281	14.90	ug/l	92
18) Methyl Acetate	4.33	43	73985	6.20	ug/l	93
20) Methylene Chloride	4.55	84	27428	1.68	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062518\
Data File : VN049517.D
Acq On : 25 Jun 2018 18:04
Operator : MD\SY
Sample : PB110537ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110537ZHE#02

Quant Time: Jun 26 02:11:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

