

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062518\
 Data File : VN049520.D
 Acq On : 25 Jun 2018 19:21
 Operator : MD\SY
 Sample : PB110537ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110537ZHE#05

Quant Time: Jun 26 02:20:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1182789	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1880988	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1539816	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	396979	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	802686	51.17	ug/l	0.00
Spiked Amount						
			Recovery	=	102.34%	
35) Dibromofluoromethane	7.59	113	696910	45.80	ug/l	0.00
Spiked Amount						
			Recovery	=	91.60%	
50) Toluene-d8	10.09	98	2557508	44.14	ug/l	0.00
Spiked Amount						
			Recovery	=	88.28%	
62) 4-Bromofluorobenzene	12.41	95	616500	32.18	ug/l	0.00
Spiked Amount						
			Recovery	=	64.36%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	57922	15.21	ug/l	99
18) Methyl Acetate	4.33	43	75674	6.52	ug/l	92
20) Methylene Chloride	4.55	84	32512	2.04	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062518\
Data File : VN049520.D
Acq On : 25 Jun 2018 19:21
Operator : MD\SY
Sample : PB110537ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110537ZHE#05

Quant Time: Jun 26 02:20:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

