

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062518\
 Data File : VN049523.D
 Acq On : 25 Jun 2018 20:38
 Operator : MD\SY
 Sample : PB110537ZHE#08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110537ZHE#08

Quant Time: Jun 26 02:27:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1111725	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1773987	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1451903	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	375348	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	755161	51.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.42%	
35) Dibromofluoromethane	7.59	113	674701	47.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.02%	
50) Toluene-d8	10.09	98	2387217	43.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.36%	
62) 4-Bromofluorobenzene	12.41	95	582696	32.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.50%	
Target Compounds						
16) Acetone	3.82	43	53167	14.85	ug/l	92
18) Methyl Acetate	4.33	43	85879	7.87	ug/l	96
20) Methylene Chloride	4.55	84	36389	2.43	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062518\
Data File : VN049523.D
Acq On : 25 Jun 2018 20:38
Operator : MD\SY
Sample : PB110537ZHE#08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110537ZHE#08

Quant Time: Jun 26 02:27:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

