

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062518\
 Data File : VN049524.D
 Acq On : 25 Jun 2018 21:04
 Operator : MD\SY
 Sample : PB110537ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110537ZHE#09

Quant Time: Jun 26 02:30:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1134819	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1825822	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1481901	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	382297	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.03	65	774721	51.47	ug/l	0.00
Spiked Amount			Recovery	=	102.94%	
35) Dibromofluoromethane	7.59	113	686938	46.50	ug/l	0.00
Spiked Amount			Recovery	=	93.00%	
50) Toluene-d8	10.09	98	2454730	43.64	ug/l	0.00
Spiked Amount			Recovery	=	87.28%	
62) 4-Bromofluorobenzene	12.41	95	598494	32.18	ug/l	0.00
Spiked Amount			Recovery	=	64.36%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.82	43	55950	15.31	ug/l	94
18) Methyl Acetate	4.33	43	95184	8.54	ug/l	98
20) Methylene Chloride	4.55	84	35312	2.31	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062518\
Data File : VN049524.D
Acq On : 25 Jun 2018 21:04
Operator : MD\SY
Sample : PB110537ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110537ZHE#09

Quant Time: Jun 26 02:30:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

