

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062518\
 Data File : VN049525.D
 Acq On : 25 Jun 2018 21:30
 Operator : MD\SY
 Sample : PB110537ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110537ZHE#10

Quant Time: Jun 26 03:09:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1084858	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1741081	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1364794	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	324749	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	728018	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
35) Dibromofluoromethane	7.59	113	648503	46.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.08%	
50) Toluene-d8	10.09	98	2306623	43.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.02%	
62) 4-Bromofluorobenzene	12.41	95	517520	29.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	58.36%	
Target Compounds						
16) Acetone	3.82	43	56245	16.10	ug/l	# 85
18) Methyl Acetate	4.33	43	87444	8.21	ug/l	94
20) Methylene Chloride	4.55	84	35816	2.45	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062518\
Data File : VN049525.D
Acq On : 25 Jun 2018 21:30
Operator : MD\SY
Sample : PB110537ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110537ZHE#10

Quant Time: Jun 26 03:09:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

