

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN062518\
 Data File : VN049535.D
 Acq On : 26 Jun 2018 2:37
 Operator : MD\SY
 Sample : PB110537ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB110537ZHE#12

Quant Time: Jun 26 03:48:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 16 01:02:01 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1193217	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1817328	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1505227	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	375995	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	780723	49.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.66%	
35) Dibromofluoromethane	7.59	113	692890	47.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.26%	
50) Toluene-d8	10.09	98	2522798	45.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.12%	
62) 4-Bromofluorobenzene	12.40	95	596379	32.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	64.44%	
Target Compounds						
16) Acetone	3.82	43	53936	14.04	ug/l	96
18) Methyl Acetate	4.34	43	76270	6.51	ug/l	92
20) Methylene Chloride	4.55	84	36659	2.28	ug/l	90
95) Naphthalene	15.14	128	16743	3.68	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN062518\
Data File : VN049535.D
Acq On : 26 Jun 2018 2:37
Operator : MD\SY
Sample : PB110537ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB110537ZHE#12

Quant Time: Jun 26 03:48:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N061618W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 16 01:02:01 2018
Response via : Initial Calibration

